ZERTIELKAT . PERTIELP

UCB_F_12.02 2012-02



CERTIFICATE

No. U8 16 04 95104 002

Holder of Certificate: Ocean Optics Inc.

4301 Metric Drive Winter Park FL 32792

USA

Production Facility(ies):

95104

Certification Mark:



Product: Electrical equ. for measurement, control and laboratory use

Integrated Sampling System

Model(s): FLAME-DA-CUV-UV-VIS, USB-ISS-UV-VIS-2,

CHEMUSB4-UV-VIS, USB-650-UV-VIS

Parameters: Rated Input Voltage: 5 V DC

Rated Input Current: 2.0 A
Degree of Protection: IPX0

Tested CAN/CSA C22 2 No. 61010-1:2012

according to: UL 61010-1:2012 EN 61010-1:2010

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in anyway. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC Guide 67. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.: 72114281-100

Date, 2016-04-21

Page 1 of 1



